

University of Technology, Sydney  
Faculty of Science

**Developing Computer Controlled Laser  
Systems for Double Blind Acupuncture Trials**

**Cameron Rogers**

Thesis submitted for the degree of

**Master of Science**

2012

## Certificate of Authorship and Originality

I certify that this thesis has not been previously submitted for a degree nor has it been submitted as part of requirements for a degree except as fully acknowledged within the text.

I also verify that the thesis has been written by me. Any help that I have received in my research work and the preparation of the thesis itself has been acknowledged.

In addition, I verify that all information sources and literature used are indicated in the thesis.

Cameron Rogers

Production Note:

Signature removed prior to publication.

Signature of Candidate

## Acknowledgement

Writing this thesis has been a long project, longer than I had originally expected, as with life things don't always go as planned. The loss of my greatest mentor, my mother Prof. Carole Rogers occurred during this project. Losing Carole was a severe loss in more ways than one, not just for me but for the University and the Acupuncture profession as a whole. My promise to Carole to see this out became one of my major motivations to keep going, even though without her the task seemed impossible at times. Therefore I dedicate this work to her memory.

My sincere thanks also has to go in no small part to my supervisor Dr. Peter Meier, as without Peter's never-ending help, work and guidance this thesis could never have seen the light of day. Whenever I was ready to throw in the towel, Peter was always there to keep me on track and focused on the next hurdle. No matter if it was the Therapeutic Goods Administration, Ethics Committee or presentations, he always gave good advice and cut through all the red tape. I often began to think it should be a higher degree in bureaucracy and Peter should be the one to get it rather than me.

Many thanks also go to Mary Garvey, who took on the role of my secondary supervisor. Mary was always supportive of my work on this research, happy to share her insights and advice.

Not least, I would like to thank Nancy Morgan who helped keep me on track and was always ready to point out what "Carole would have wanted me to do", when I lost my path. Nancy often offered help when I was feeling overwhelmed.

My acknowledgements would not be complete without mentioning Melissa Miller. Melissa was always happy to critically read my endless scribbled work and listen to my presentation even if the topic of laser research probably bored her half to death. She was the voice of reason when I would assume too much prior knowledge on the part of the audience. Melissa always encouraged me to see the whole thing through to the end. With the heart of a lioness, she was by my side through thick and thin.

## Contents

Certificate of Authorship and Originality .....	II
Acknowledgement.....	III
List of Tables.....	VI
List of Figures .....	VI
Abstract .....	VII
Chapter 1 Introduction .....	1
1.1 Background to the Study .....	1
1.2 What is Laser?.....	2
1.3 LLLT .....	3
1.4 LLLT and Laser Acupuncture.....	4
1.5 Laser as a Substitute for Needles .....	5
1.6 Laser Research Delimiters .....	7
1.7 Aims of This Study .....	7
1.8 Objectives of This Study.....	8
1.9 Methods.....	8
1.10 Benefits of This Study.....	9
Chapter 2 Literature Review .....	10
2.1 Wavelength .....	10
2.1.1 Violet 405 nm.....	15
2.1.2 Visible 405 nm to 685 nm.....	15
2.1.3 Non-visible Infrared (IR) 700 nm to 10.6 um .....	16
2.2 Blinding Analysis.....	17
2.3 Dosage.....	20
2.4 Time .....	21
2.5 Laser Output Power.....	21
2.6 Randomisation .....	23
2.7 Delivery System.....	25
2.8 Monitoring and Recording of Intervention .....	26
2.9 Summary .....	26
Chapter 3 Design.....	27
3.1 Developing a Computer Controlled Laser System for Double Blind Studies	27
3.2 Design Parameters.....	27
3.2.1 Wavelength .....	27
3.2.2 Blinding.....	28
3.2.3 Dosage.....	29
3.2.4 Time .....	29
3.2.5 Power .....	30
3.2.6 Randomisation .....	31

3.2.7	Delivery System.....	31
3.2.8	Monitoring and Recording.....	33
3.3	The System and Computer Interface.....	34
3.4	Testing.....	37
3.5	Procedure.....	38
3.6	Summary of Design .....	38
Chapter 4 Results .....		39
4.1	Outputs From the Laser System.....	39
4.2	Questionnaire Results.....	42
4.3	Summary .....	43
Chapter 5 Conclusion.....		44
5.1	Conclusions Drawn from Literature Review .....	44
5.2	The Design Process .....	44
5.3	Results .....	45
5.3.1	Limitations of the Study.....	45
5.3.2	Strengths of the Study .....	46
5.3.3	Summary .....	47
Appendices.....		48
Appendix A-1.....		48
	Laser Controller Operating Procedure Instructions .....	48
Appendix A-2.....		49
	Programme Operation Instructions .....	49
Appendix A-3.....		52
	Circuit Function Description.....	52
Appendix A-4.....		53
	Circuit Diagram of Laser Unit .....	53
Appendix B-1 .....		54
	Glossary .....	54
Appendix B-2.....		60
	Example Participant Survey Form: .....	60
	Example Practitioner Survey Form:.....	61
References.....		62

## **List of Tables**

Table 1: Visible wavelengths and unrecorded by number of reviewed papers.....	12
Table 1: (cont): Infrared wavelengths by number of reviewed papers .....	13
Table 2: Comparison of wavelengths used in papers reviewed and response .....	14
Table 3: Blinding strategies as by papers reviewed .....	18
Table 4: Range of dosage delivered by study and the related clinical outcome: .....	20
Table 5: Power ranges utilised in the studies under review .....	22
Table 6: Methods of randomisation in the papers reviewed .....	24
Table 7: Operator rating of the system.....	42
Table 8: Participant ratings of the system .....	42

## **List of Figures**

Figure 1: Photo of moxibustion .....	6
Figure 2: The visible spectrum.....	10
Figure 3: Drawing of the fundamental system design .....	32
Figure 4: Drawing of the collimated laser output .....	32
Figure 5: Drawing of the detection of the laser activation.....	33
Figure 6: Photo of the prototype laser unit.....	35
Figure 7: Screenshot of computer interface .....	36
Figure 8: Point location of Pericardium 6 .....	37
Figure 9: Illustrated screen dump of laser programme .....	48
Figure 10: Circuit diagram of the laser computer interface .....	53

## Abstract

Title: Developing Computer Controlled Laser Systems for Double Blind Acupuncture Trials

Background: Double blind clinical testing is considered critical for advancing most areas of medical research. Much of the acupuncture laser and Low Level Laser Therapy (LLLT) research lacks double blinding and suffers from poor trial design. While some research studies have shown a potential increased clinical effect in visible wavelength laser trials visible wavelength lasers are generally overlooked in research, due to the difficulties with double blinding.

Objectives: This study sought to design, construct and test a system that could randomise, control, monitor and record the delivery of visible laser in real-time, and allow for double blinding.

Method: The study utilised a novel approach to incorporate computer controlled lasers within opaque hoods. An optical sensor feedback was incorporated to objectively measure the delivery of laser in a double blinded randomised controlled trial (RCT) situation.

The equipment was designed and tested for future researchers to operate with a minimum of training.

The initial testing was designed around a *proof of concept* study using the equipment to stimulate Pericardium 6 (PC 6) with a visible (635 nm) 10 mW laser while monitoring changes in Heart Rate Variability (HRV). We only sought to trial the equipment under laboratory conditions to see if a double blind condition could be achieved. It was beyond the scope of this study to undertake a RCT of the effects of PC 6 on HRV.

Results: The equipment was successful in maintaining double blind conditions. Laser was delivered and monitored in real-time, the operational parameters were recorded and test participants reported no discomfort associated with the use of the equipment. The

computer successfully controlled the random delivery of laser. The test participants and operators were at no time aware of the state of the laser. The operators reported no difficulties in using the unit.

Conclusion: A laser research tool was successfully created and the study demonstrated the system's unique ability to overcome all the difficulties identified for visible laser wavelength double blind research.

The system is currently designed with two laser hoods to allow for bilateral acupuncture point testing. Future improvements could include additional laser hoods for multiple acupuncture point treatments. Construction of alternate hoods would allow for testing of different wavelengths including infrared and different laser power levels.

The system combines double blind testing with physiological monitoring, which allows researchers to record the physiological effects that accompany laser therapies.