

**NON-INVASIVE DETECTION OF
HYPERGLYCAEMIA IN TYPE 1 DIABETIC
PATIENTS USING ELECTROCARDIOGRAPHIC
SIGNALS**

**By
Linh Lan Nguyen**

A thesis submitted in partial fulfilment of the requirements for the
Degree of Doctor of Philosophy



**University of Technology Sydney
Faculty of Engineering and Information Technology**

2014

CERTIFICATE OF AUTHORSHIP/ORIGINALITY

I, Linh Lan Nguyen, certify that the work in the thesis has not previously been submitted for a degree nor has it been submitted as part of requirements for a degree except as fully acknowledged within the text.

I also certify that this thesis has been written by me. Any help that I have received in my research work and the preparation of the thesis itself has been acknowledged. In addition, I certify that all information sources and literature used are indicated in the thesis.

Signature of Candidate



*This thesis is especially dedicated to my dearest family
for their endless love, care and encouragement ...*

Acknowledgements

Firstly of all, I would like to thank Vietnamese Ministry of Education and Training (MOET) and University of Technology Sydney (UTS) for granting me a University of Technology Sydney – Vietnam International Education Development scholarship (UTS-VIED) without which I would not be able to carry out this research. I also would like to thank Hanoi University of Science and Technology for giving me an opportunity to attend the PhD program.

My great appreciation goes to my principal supervisor Professor Hung Nguyen for his great intellectual guidance and supervision during my PhD candidature. He gave me the opportunity to perform research work in a very practical and interesting scientific area. His advice and support have contributed enormously to the quality of my research and have made my PhD challenging, inspiring and enjoyable.

I would like to extend my gratitude to my co-supervisor Dr Steven Su for his consultation and encouragement. I am also extremely grateful to Lesley Nguyen for her enthusiasm in proofreading and correcting my thesis. I would like to sincerely thank Dr Nejhddeh Ghevondian for his thoughtful advice and ideas throughout the time I carried out this research.

I appreciate all my colleagues and staff in Centre for Health Technologies who supported me in my study with encouragement and confidence during the past few years.

Last but not least, I would like to express my heartfelt gratitude to my parents, my brother Hoang and especially my husband Dai and my son Phong for their great love, understanding and support over the years that enabled me to complete this thesis.

Table of contents

CHAPTER 1. INTRODUCTION	1
1.1 BACKGROUND	2
1.2 MOTIVATION OF THE THESIS	6
1.3 AIMS OF THESIS	8
1.4 CONTRIBUTIONS	10
1.5 STRUCTURE OF THESIS	11
1.6 PUBLICATIONS RELATED TO THESIS	13
CHAPTER 2. LITERATURE REVIEW	15
2.1 HYPERGLYCAEMIA	15
2.1.1 Definition and causes of hyperglycaemia	15
2.1.2 Blood glucose regulation.....	16
2.1.3 Blood glucose threshold definition	18
2.2 EFFECT OF HYPERGLYCAEMIA ON THE ELECTRICAL ACTIVITIES OF THE HEART.....	19
2.2.1 Effect of hyperglycaemia on electrocardiogram intervals	19
2.2.2 Effect of hyperglycaemia on Heart rate variability	23
2.3 CURRENT RESEARCH OF HYPERGLYCAEMIA DETECTION.....	28
2.3.1 Test of blood glucose level - Finger stick method	29
2.3.2 Continuous glucose monitoring system (CGMS)	31
2.3.3 Reverse iontophoresis technique.....	33

2.3.4	Near-infrared light (NIR) spectroscopy technique.....	33
2.3.5	Early detection of β cell death.....	35
2.3.6	Testing of methyl nitrates in exhaled air.....	35
2.3.7	Measuring glucose concentration in tears	36
2.4	THE GAP IN LITERATURE REVIEW.....	38
2.5	PROPOSED STRATEGY OF HYPERGLYCAEMIA DETECTION	40
CHAPTER 3. ELECTROCARDIOGRAPHIC BASED HYPERGLYCAEMIA DETECTION EMPLOYING ARTIFICIAL NEURAL NETWORK		
3.1	INTRODUCTION.....	42
3.2	ARTIFICIAL NEURAL NETWORK FOR HYPERGLYCAEMIA DETECTION	44
3.2.1	ECG acquisition	44
3.2.2	ECG feature extraction.....	46
3.2.3	ANN classifier for hyperglycaemia detection.....	53
3.3	EXPERIMENTAL RESULTS	62
3.3.1	ECG parameters obtained from the study	62
3.3.2	Performance of hyperglycaemia detection using ANN	66
3.3.3	Comparison with other methods	70
3.4	DISCUSSION	71
CHAPTER 4. COMBINED PRINCIPAL COMPONENT ANALYSIS AND NEURAL NETWORK FOR HYPERGLYCAEMIA DETECTION		
4.1	INTRODUCTION.....	76
4.2	PRINCIPAL COMPONENT ANALYSIS (PCA)	78
4.2.1	Overview of PCA.....	78
4.2.2	Rotation.....	80

4.2.3	Analysis of principle component	81
4.3	TWO-STAGE ANN CLASSIFIER USING PCA FOR HYPERGLYCAEMIA DETECTION	82
4.3.1	First stage – PCA for feature reduction.....	83
4.3.2	Second stage – ANN classifier.....	85
4.4	EXPERIMENTAL RESULTS	86
4.4.1	PCA results	86
4.4.2	Subgroup data set	89
4.4.3	Performance of two-stage ANN classifier using PCA for hyperglycaemia detection.....	91
4.5	DISCUSSION	94
CHAPTER 5. PARTICLE SWARM OPTIMISATION BASED NEURAL NETWORK FOR HYPERGLYCAEMIA DETECTION		98
5.1	INTRODUCTION.....	98
5.2	PARTICLE SWARM OPTIMISATIONBASED NEURAL NETWORK (PSO-NN).....	100
5.2.1	PSO algorithm.....	100
5.2.2	Adaptation of PSO to train neural networks	100
5.2.3	Hyperglycaemia detection using PSO-NN.....	103
5.3	EXPERIMENTAL RESULTS	107
5.3.1	Performance of PSO-NN.....	107
5.3.2	Comparison of PSO-NN with other NN-based algorithms.....	116
5.4	DISCUSSION	118
CHAPTER 6. DISCUSSION, CONCLUSIONS AND FUTURE DIRECTIONS.....		122
6.1	DISCUSSION	122

6.1.1	Hyperglycaemia and abnormalities in ECG signals	122
6.1.2	Comparison of experimental results.....	125
6.2	CONCLUSIONS	128
6.2.1	Summary and conclusions	128
6.2.2	Main findings	131
6.3	FUTURE DIRECTIONS.....	132
APPENDIX A. ARTIFICIAL NEURAL NETWORK		135
APPENDIX B. PARTICLE SWARM OPTIMIZATION		143
APPENDIX C. LABVIEW PROGRAM.....		149
APPENDIX D. MATLAB CODE		151
REFERENCES.....		164

List of figures

Figure 2.1: Blood glucose curves of three glycaemic states after a meal containing 100 grams of glucose	17
Figure 2.2: QT , PR and RR intervals in an electrocardiogram.....	20
Figure 2.3: QT_C interval changes during 0, 60, 120 minutes of hyperglycaemia vs. normoglycaemia in study of Gordin et al. (Gordin <i>et al.</i> , 2008).....	21
Figure 2.4: Individual correlation QT_C per hour versus glycaemia ($r = 0.672$, $p = 0.003$) in a study of Suys et al. (Suys <i>et al.</i> , 2006).....	22
Figure 2.5: PR interval during hyperglycaemic clamps in Marfella <i>et al.</i> 's study (Marfella <i>et al.</i> , 2000). Left dark bar basal values, right light bar 120 min clamp values.....	23
Figure 2.6: Overview of technologies for blood glucose monitoring (Ferrante do Amaral et al., 2008).....	29
Figure 2.7: Reference glucose and NIR-predicted glucose profiles (Malin <i>et al.</i> , 1999)	34
Figure 2.8: Plasma glucose and exhaled methyl nitrate profiles in one child with T1DM in study of Novak et al. (2007).....	36
Figure 2.9: Contact based-lens sensors for tear glucose monitoring developed by researchers at University of Washington (Parviz, 2009)	37
Figure 2.10: Measured glucose response and calibration curve of sensor in study of Yao et al. (2011).....	38
Figure 2.11: ECG-based hyperglycaemia detection system using ANN	41
Figure 3.1: General structure of hyperglycaemia detection based on ANN and ECG parameters	44

Figure 3.2: Actual blood glucose level profiles in 10 T1DM patients.....	46
Figure 3.3: Block diagram of ECG pre-processing.....	47
Figure 3.4: Schematic representation of normal ECG points and intervals.....	48
Figure 3.5: The ECG signals with denoted significant points, upper graph with non-hyper ECG and lower graph with hyper ECG	50
Figure 3.6: Processed ECG signals with zero crossing detector and calculated ECG parameters	51
Figure 3.7: HRV analysis results in Kubios for 1 patient under hyperglycaemic condition.....	53
Figure 3.8: Neural network structure	54
Figure 3.9: Hyperglycaemia detection feed-forward Multilayer neural network classifier and input of ECG parameters	56
Figure 3.10: Training and validation error cycles of hyperglycaemia classification using ANN	67
Figure 3.11: ROC curve: Sensitivity vs. 1-Specificity.....	68
Figure 4.1: Hyperglycaemia detection using two-stage PCA-Multilayer neural network classifier and input of ECG parameters	83
Figure 4.2: ROC plot.....	92
Figure 5.1: PSO based feed-forward Neural network for hyperglycaemia detection	103
Figure 5.2: Swarm motions on the 3-D coloured contour maps	109
Figure 5.3: History graph of global best positions.....	110
Figure 5.4: The function value of the global best obtained with 1000 iterations... ..	110
Figure 5.5: The function value of the global best obtained with 1200 iterations... ..	111
Figure 5.6: Performance curves of PSO-NN at various velocities.....	112
Figure 5.7: Variations of the weights of PSO-NN at various velocities:.....	112

Figure 5.8: PSO-NN Performance curves at various acceleration constants	113
Figure 5.9: Values of weights in PSO-NN at $c_1 = c_2 = 1$ and $c_1 = c_2 = 2$	114
Figure 5.10: The ROC curve of best performance of PSO-NN	116

List of tables

Table 2.1: Selected time and frequency domain measures of HRV	26
Table 3.1: Changes of natural 16 ECG parameters under hyperglycaemic conditions compared to normoglycaemic conditions	64
Table 3.2: Changes of normalized 16 ECG parameters under hyperglycaemic conditions compared to normoglycaemic conditions.....	65
Table 3.3: Mean values of training, validation and testing results using five different algorithms.....	69
Table 3.4: Best performance of hyperglycaemia detection using 16 ECG parameters and five different algorithms.....	69
Table 3.5: Comparison of best performances of three different classification techniques for hyperglycaemia detection.....	71
Table 4.1: Collinearity Statistics of original ECG data.....	86
Table 4.2: Pearson correlations matrix of original ECG parameters	87
Table 4.3: KMO and Bartlett's test for ECG data.....	88
Table 4.4: Total variance and rotated principal components loadings.....	90
Table 4.5: Performances of ECG classifier using PCA and LM algorithm	92
Table 4.6: Testing performances of neural network using PCA and five different algorithms.....	93
Table 4.7: Testing performances of neural network without PCA using all 16 original parameters and five different algorithms	94
Table 5.1: ANN parameters used in the PSO-NN model.....	105
Table 5.2: PSO parameters used in the PSO-NN model.....	105
Table 5.3: Performances of hyperglycaemia detection at different hidden nodes .	115
Table 5.4: Mean values of training, validation and testing results over 20 runs ...	116
Table 5.5: Training execution errors of PCA-LM-NN and PSO-NN models.....	117

Table 5.6: Cohen’s Kappa statistic of PCA-LM-NN and PSO-NN models	117
Table 5.7: Best testing results of three methods	118
Table 6.1: Best performances of Training, Validation and Testing sets of three proposed methods	125
Table 6.2: Comparisons of three proposed methodologies.....	127

List of Abbreviations

ANN	:	Artificial neural network
AUC	:	Area under the curve
BGL	:	Blood glucose level
CGF	:	Conjugate gradient back propagation with Fletcher-Reeves updates
CGMS	:	Continuous glucose monitoring system
ECG	:	Electrocardiography
FFT	:	Fast Fourier transform
GA	:	Genetic algorithm
GDX	:	Gradient descent with momentum
<i>gm</i>	:	Geometric mean
HRV	:	Heart rate variability
KNN	:	K-nearest neighbour
LDA	:	Linear Discriminant Analysis
LM	:	Levenberg Marquardt
LM-NN	:	Neural network using Levenberg Marquardt algorithm
MSE	:	Mean squared error
NIR	:	Near-infrared light
NN	:	Neural network
PC	:	Principle component
PCA	:	Principal component analysis
PCA-CGF	:	Neural network using principal component analysis and Conjugate gradient backpropagation with Fletcher-Reeves updates algorithm
PCA-GDX	:	Neural network using principal component analysis and Gradient descent with momentum algorithm
PCA-LM	:	Neural network using principal component analysis and Levenberg Marquardt algorithm
PCA-RP	:	Neural network using principal component analysis and Resilient backpropagation algorithm
PCA-SCG	:	Neural network using principal component analysis and Scaled conjugate gradient algorithm
PSO	:	Particle swarm optimisation

PSO-NN	:	Neural network using particle swarm optimisation
ROC	:	Receiver operating characteristic
RP	:	Resilient backpropagation
SCG	:	Scaled conjugate gradient
<i>Sen</i>	:	Sensitivity
<i>Spec</i>	:	Specificity
T1DM	:	Type 1 Diabetes mellitus
VIF	:	Variance inflation factor

List of Symbols

Q	:	the beginning of QRS complex
R	:	the peak of QRS complex
S	:	the end of QRS complex
P	:	the beginning of P -wave
T_p	:	the peak of T -wave
T_E	:	the end of T -wave
RR	:	the interval between two consecutive R points
PR	:	the interval from the beginning of P -wave to the beginning of QRS complex
QT	:	the interval from Q point to the end of T -wave (T_E)
QT_C	:	the corrected QT interval by heart rate (Bazett's formula)
RT	:	the interval from R point to the end of T -wave (T_E)
RT_C	:	the corrected RT interval by heart rate (Bazett's formula)
T_pT_E	:	the interval from the peak of T -wave (T_p) to the end of T -wave (T_E)
T_pT_{EC}	:	the corrected T_pT_E interval by heart rate (Bazett's formula)
$MeanRR$:	mean RR interval
$SDNN$:	the standard deviation of the RR interval index
$RMSSD$:	the root mean square of successive RR interval differences
$pNN50$:	the percentage of beats with a consecutive RR interval more than 50 ms
$HRVi$:	HRV triangular index
$TINN$:	baseline width of the RR interval histogram evaluated through triangular interpolation
VLf	:	Very low frequency
LF	:	Low frequency
HF	:	High frequency
$TotalPw$:	Total spectral power
LF/HF	:	ratio between LF and HF components
p_{best}	:	best solution of the particle has achieved so far
g_{best}	:	the global best
l_{best}	:	the local best
w_i	:	position of the particle
v_i	:	velocity of the particle

c_1, c_2	:	acceleration constants
ω	:	inertia factor
ω_{max}	:	initial weight
ω_{min}	:	final weight
v_{max}	:	maximum allowable velocity
$v_{max-perc}$:	velocity clamping percentage

Abstract

Hyperglycaemia is the medical term for a state caused by a high level of blood glucose, resulting from defects in insulin secretion, insulin action, or both. Hyperglycaemia is a common dangerous complication to glycaemic control in Type 1 diabetic patients. The chronic hyperglycaemia of diabetes is associated with long-term damage, dysfunction, and failure of different organs, especially the eyes, kidneys, nerves, heart, and blood vessels. Therefore, reliable detection of hyperglycaemic episodes is important in order to avoid major health conditions.

Conventionally, diabetic patients need to frequently monitor blood glucose levels to determine whether they have hyperglycaemia or not. A patient has to prick their finger (finger-stick) for a drop of blood several times a day, which can therefore significantly discourage many patients from periodically checking blood glucose levels. Another choice for hyperglycaemia detection might be continuous glucose monitoring systems (CGMS), which measure the glucose level in the interstitial fluid. For patients using CGMS, finger-sticks are still required to calibrate the sensor. The main shortcoming of CGMS is that glucose levels in interstitial fluid lag temporally behind blood glucose values, normally 10-15 minutes, which absolutely limits the accuracy of the detection. There is a strong demand to have a non-invasive technique to help patients to diagnose the disease easily and painlessly. Few methods have been reported to detect hyperglycaemia non-invasively or minimally invasively such as in exhaled methyl nitrates, and early detection of ongoing β cell death. However, the purpose of these studies was on real-time glucose control rather than disease diagnosis.

Electrocardiography (ECG) is a broadly used technique to obtain a quick, non-invasive clinical and research screen for diagnosing abnormal rhythms of the heart caused by diseases. In fact, observations of ECG changes have been found in hypoglycaemia and hyperglycaemia states in T1DM, such as increased heart rate and prolongation of QT interval in hypoglycaemia, whereas hyperglycaemia was related to reduced heart rate variability. By using these findings in hypoglycaemia,

researchers have developed an effective and sensitive system to detect hypoglycaemia non-invasively. These excellent performances of hypoglycaemia detection using ECG is the motivation of this thesis to study the effect of hyperglycaemia on ECG signals, and based on the findings to exploit the computational intelligence on the non-invasive detection of hyperglycaemia.

This research firstly explores the changes of ECG parameters associated with the hyperglycaemic state in T1DM. The ECG parameters consist of ECG intervals relating to repolarisation phase and heart rate variability (HRV) measures. A clinical study of ten T1DM patients and ECG feature extraction process are conducted to collect ECG features. Statistical analysis is then applied to every ECG feature to estimate the significant difference between hyperglycaemic and normoglycaemic states. The results show that the selected ECG parameters in hyperglycaemia differ significantly from those in normoglycaemia ($p < 0.05$). It implies that certain ECG parameters are correlated with high blood glucose levels and they possibly contribute to the performance of hyperglycaemia detection. Thus, the ECG parameters are used for input data of hyperglycaemia classifiers in this thesis.

Furthermore, the thesis introduces novel computational intelligent methods for hyperglycaemia detection using the ECG parameters. A neural network using Levenberg-Marquardt algorithm is the first method explored for hyperglycaemia detection in this thesis, known as LM-NN. The second algorithm is the integration of principal component analysis (PCA) with a neural network utilising the Levenberg-Marquardt algorithm, which is called a PCA-LM-NN network. PCA is a useful tool for dimensionality reduction to diminish the computational requirement and overcome the problem of multicollinearity. It is employed to filter the data so that only the significant independent ECG variables responsible for the high blood glucose levels can be used as input for the network training, in order that the neural network performs well for hyperglycaemia detection. The third method is for the improvement of the second method where particle swarm optimization is included. This algorithm is a combination of PCA, PSO and neural network, which is called PSO-NN. The PSO is utilised as an effective training algorithm to optimise the weights of the neural network. The proposed methods are compared with each other

and with other traditional classifiers. All the algorithms are investigated with the clinical electrocardiographic data extracted from ten T1DM patients.

The results show that the performance of PCA-LM model for hyperglycaemia detection is better than that of LM-NN (70.88% vs. 67.94%, in terms of geometric mean). In addition, the PSO-NN outperforms the PCA-LM-NN (77.58% vs. 70.88%, in terms of geometric mean). In short, the PSO-NN significantly improves the performances of both the LM-NN and PCA-LM-NN, with considerable sensitivity, specificity and geometric mean of 82.35%, 73.08% and 77.58%, respectively.